AM-2
Acupuncture Training Pad
(With indicator light and chime)

SAKAMOTO MODEL
CORPORATION

User's Manual (131030)

－ＩＮＤＥＸ－

●Introduction/Table of Contents 1
●Purpose/Content 2
●Cautions on Use/Storage Instructions 3
●Instructions for Use 4

! Note	Confirm the main unit and accessories directly after the product arrives. If there is any defect or breakage, inform the dealer or Sakamoto Model.

株式会社坂本モデル

Introduction

Thank you for purchasing our product,

“Acupuncture Training Pad (with indicator light and chime).”

Read this user’s manual carefully in order to use the product for a long time appropriately.

This product is closely checked before factory shipment. However, when this product arrives, open it soon, and check if anything on the main unit and accessories are missing or broken.

Table of contents

Introduction／Table of Contents…… 1

Purpose／Content…….. 2

Cautions on Use／Storage Instructions…… 3

Instructions for Use…… 4

Purpose

Acupuncture training to practice using the
“dominant hand to hold the needle,”
“non-dominant hand to assist insertion,”
“perpendicular insertion” and “oblique insertion”
Shallow and deep target points
Equipped with an indicator light and chime to evaluate correct insertion

Content

1) Acupuncture training pad – Qty. 1
2) Training belt – Qty. 1
3) Rubber band – Qty. 1
4) Batteries (AA alkaline batteries) – Qty. 2



Cautions on Use

1) Use new needle tips when performing acupuncture training with this product. Otherwise, it can cause the skin of the pad or the sensor to deteriorate.

2) Take extra care when handling the product because it is fragile and will break if dropped.

3) If the product gets dirty, use a neutral detergent on a soft cloth to remove the dirt or stain. (Do not use solvents such as benzine, thinner or polishing powder.)

4) Use the two AA size alkaline batteries.
Make sure that the positive and negative sides on the batteries are inserted correctly. Replace the batteries with new ones when the chime sound becomes weak or when the indicator light fades.

5) If a defect or problem with the product is discovered during use, stop using the product immediately and contact Sakamoto Model or an authorized dealer.

6) This product is a simulator capable of supporting advanced medical training. However, Sakamoto Model does not guarantee the same results during actual treatment.

7) Please note that the specifications of this product may be changed without prior notice in order to make improvements to the product.

Storage Instructions

1) Avoid storing the product in direct sunlight or in locations with high temperature and humidity.

2) Always remove the batteries when storing the product.
Be careful when storing because if the batteries are not removed during storage, the batteries could corrode and disrupt the power supply.

Instructions for Use

1) Insert the two AA alkaline batteries into the battery compartment and ensure that they are inserted correctly.

2) Turn the power switch on.

3) This product has two depths. Look at the display on the side, select a shallow or deep target point and begin practicing.

4) If the point of insertion is correct, the green indicator light turns on and a chime sounds.

5) The rubber band that is provided can be attached to the hooks on the sides of the pad in order to help locate the target point.

SAKAMOTO MODEL
CORPORATION

Head office
34,Higashitakagicho,Shimogamo
Sakyo-ku, KYOTO 606-0865, JAPAN
Phone: +81-75-701-1135
Fax: +81-75-722-5638
E-mail:info@sakamoto-model.com
URL:http://sakamoto-model.co.jp

Overseas Sales Department
1-9-29,Senbahigashi,Minoh City,Osaka
562-0035,JAPAN
Phone: +81-72-729-8350
Fax: +81-72-729-8385
E-mail:trade@sakamoto-model.com